

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061725\
 Data File : VN087047.D
 Acq On : 17 Jun 2025 10:20
 Operator : JC\MD
 Sample : VN0617WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0617WBL01

Quant Time: Jun 18 01:43:00 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 02:12:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	200915	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	403651	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	368689	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	170553	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	148580	55.234	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	110.460%
35) Dibromofluoromethane	8.177	113	126383	52.833	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.660%
50) Toluene-d8	10.565	98	504769	53.303	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	106.600%
62) 4-Bromofluorobenzene	12.847	95	176121	50.058	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	100.120%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061725\
Data File : VN087047.D
Acq On : 17 Jun 2025 10:20
Operator : JC\MD
Sample : VN0617WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0617WBL01

Quant Time: Jun 18 01:43:00 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 07 02:12:50 2025
Response via : Initial Calibration

