

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061725\
 Data File : VN087053.D
 Acq On : 17 Jun 2025 12:56
 Operator : JC\MD
 Sample : Q2299-16DL 5X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE134D4-20250610DL

Quant Time: Jun 18 01:46:41 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 02:12:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	272313	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	524978	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	453855	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	219879	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	169747	46.557	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	93.120%		
35) Dibromofluoromethane	8.177	113	153523	49.346	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	98.700%		
50) Toluene-d8	10.565	98	615838	50.002	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	100.000%		
62) 4-Bromofluorobenzene	12.847	95	221456	48.397	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	96.800%		
Target Compounds						
9) 1,1,2-Trichlorotrifluo...	4.406	101	192795	64.940	ug/l	98
44) Trichloroethene	9.353	130	21017	5.847	ug/l	96
64) Tetrachloroethene	11.106	164	16088	5.601	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061725\
Data File : VN087053.D
Acq On : 17 Jun 2025 12:56
Operator : JC\MD
Sample : Q2299-16DL 5X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE134D4-20250610DL

Quant Time: Jun 18 01:46:41 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 07 02:12:50 2025
Response via : Initial Calibration

