

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061725\
 Data File : VN087055.D
 Acq On : 17 Jun 2025 13:39
 Operator : JC\MD
 Sample : Q2299-17DL 10X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE134D3-20250610DL

Quant Time: Jun 18 01:47:28 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 02:12:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	309321	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	598780	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	533505	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	255285	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	198501	47.930	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	95.860%
35) Dibromofluoromethane	8.171	113	176254	49.670	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.340%
50) Toluene-d8	10.571	98	729652	51.941	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.880%
62) 4-Bromofluorobenzene	12.847	95	260591	49.930	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	99.860%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	4.400	101	197130	58.456	ug/l	98
27) cis-1,2-Dichloroethene	7.506	96	1555	0.339	ug/l	81
44) Trichloroethene	9.359	130	55375	13.506	ug/l	93
64) Tetrachloroethene	11.106	164	70237	20.803	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061725\
Data File : VN087055.D
Acq On : 17 Jun 2025 13:39
Operator : JC\MD
Sample : Q2299-17DL 10X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE134D3-20250610DL

Quant Time: Jun 18 01:47:28 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 07 02:12:50 2025
Response via : Initial Calibration

