

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061725\
 Data File : VN087059.D
 Acq On : 17 Jun 2025 15:04
 Operator : JC\MD
 Sample : Q2312-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-1

Quant Time: Jun 18 01:49:08 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 02:12:50 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	209320	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	420116	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	383973	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	178960	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	149127	53.211	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 106.420%	
35) Dibromofluoromethane	8.171	113	127798	51.331	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.660%	
50) Toluene-d8	10.570	98	520900	52.851	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 105.700%	
62) 4-Bromofluorobenzene	12.847	95	185671	50.704	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 101.400%	
Target Compounds						
					Qvalue	
16) Acetone	4.447	43	35373	23.801	ug/l	95
18) Methyl Acetate	5.047	43	24744	5.713	ug/l	100
20) Methylene Chloride	5.306	84	22898	8.229	ug/l	93
37) Ethyl Acetate	7.571	43	17245	3.652	ug/l	97
43) Isopropyl Acetate	8.694	43	11145	1.470	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061725\
Data File : VN087059.D
Acq On : 17 Jun 2025 15:04
Operator : JC\MD
Sample : Q2312-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-1

Quant Time: Jun 18 01:49:08 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 07 02:12:50 2025
Response via : Initial Calibration

