

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061725\
 Data File : VN087064.D
 Acq On : 17 Jun 2025 16:51
 Operator : JC\MD
 Sample : Q2329-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TT188S1-20250612

Quant Time: Jun 18 01:51:16 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 02:12:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	262569	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	525578	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	472016	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	222915	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	182650	51.956	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	103.920%
35) Dibromofluoromethane	8.177	113	157445	50.549	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.100%
50) Toluene-d8	10.571	98	643870	52.219	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	104.440%
62) 4-Bromofluorobenzene	12.847	95	227515	49.664	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	99.320%
Target Compounds						
						Qvalue
16) Acetone	4.453	43	1951	1.047	ug/l	# 73
44) Trichloroethene	9.353	130	15452	4.294	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061725\
Data File : VN087064.D
Acq On : 17 Jun 2025 16:51
Operator : JC\MD
Sample : Q2329-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TT188S1-20250612

Quant Time: Jun 18 01:51:16 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 07 02:12:50 2025
Response via : Initial Calibration

