

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061725\
 Data File : VN087065.D
 Acq On : 17 Jun 2025 17:13
 Operator : JC\MD
 Sample : Q2329-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TT119S1-20250612

Quant Time: Jun 18 01:51:37 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 02:12:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	282994	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	562247	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.864	117	504751	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	235936	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	194771	51.405	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 102.800%	
35) Dibromofluoromethane	8.177	113	171035	51.331	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.660%	
50) Toluene-d8	10.570	98	690846	52.374	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 104.740%	
62) 4-Bromofluorobenzene	12.847	95	244256	49.841	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 99.680%	
Target Compounds						
					Qvalue	
44) Trichloroethene	9.359	130	4718	1.226	ug/l	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061725\
Data File : VN087065.D
Acq On : 17 Jun 2025 17:13
Operator : JC\MD
Sample : Q2329-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TT119S1-20250612

Quant Time: Jun 18 01:51:37 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 07 02:12:50 2025
Response via : Initial Calibration

