

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061725\
 Data File : VN087067.D
 Acq On : 17 Jun 2025 17:56
 Operator : JC\MD
 Sample : Q2329-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TT162S1-20250613

Quant Time: Jun 18 01:52:19 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 02:12:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	290343	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	583906	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	526439	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	248696	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	206924	53.230	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery	= 106.460%		
35) Dibromofluoromethane	8.177	113	176110	50.893	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery	= 101.780%		
50) Toluene-d8	10.565	98	719327	52.511	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery	= 105.020%		
62) 4-Bromofluorobenzene	12.847	95	255546	50.211	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery	= 100.420%		
Target Compounds						Qvalue
44) Trichloroethene	9.359	130	25133	6.286	ug/l	97
64) Tetrachloroethene	11.106	164	2198	0.660	ug/l	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061725\
Data File : VN087067.D
Acq On : 17 Jun 2025 17:56
Operator : JC\MD
Sample : Q2329-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TT162S1-20250613

Quant Time: Jun 18 01:52:19 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 07 02:12:50 2025
Response via : Initial Calibration

