

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061818\
 Data File : VN049334.D
 Acq On : 18 Jun 2018 11:20
 Operator : MD\SY
 Sample : VN0618WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0618WBL01

Quant Time: Jun 19 01:42:57 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:02:01 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	1637764	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	2452596	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	2056386	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	628717	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	989574	45.56	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.12%	
35) Dibromofluoromethane	7.59	113	906297	45.67	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.34%	
50) Toluene-d8	10.09	98	3425999	45.35	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.70%	
62) 4-Bromofluorobenzene	12.40	95	903574	36.17	ug/l	0.00
Spiked Amount	50.000		Recovery	=	72.34%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061818\
Data File : VN049334.D
Acq On : 18 Jun 2018 11:20
Operator : MD\SY
Sample : VN0618WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0618WBL01

Quant Time: Jun 19 01:42:57 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:02:01 2018
Response via : Initial Calibration

