

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061818\
 Data File : VN049337.D
 Acq On : 18 Jun 2018 12:38
 Operator : MD\SY
 Sample : J3493-01ME
 Misc : 5.10g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-1ME

Manual Integrations
 APPROVED

MMDadoda
 6/19/2018 1:55:41 PM

Quant Time: Jun 19 12:11:17 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:02:01 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	1278699	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1916503	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1700331	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	609567	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	792258	46.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.42%	
35) Dibromofluoromethane	7.59	113	709914	45.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.58%	
50) Toluene-d8	10.09	98	2676884	45.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.68%	
62) 4-Bromofluorobenzene	12.40	95	816008	41.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	83.60%	
Target Compounds						
27) cis-1,2-Dichloroethene	6.82	96	148840	8.625	ug/l	82
44) Trichloroethene	8.83	130	229268	12.893	ug/l	94
64) Tetrachloroethene	10.63	164	1428852	82.008	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN061818\
Data File : VN049337.D
Acq On : 18 Jun 2018 12:38
Operator : MD\SY
Sample : J3493-01ME
Misc : 5.10g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-1ME

Manual Integrations
APPROVED

MMDadoda
6/19/2018 1:55:41 PM

Quant Time: Jun 19 12:11:17 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:02:01 2018
Response via : Initial Calibration

