

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061818\
 Data File : VN049340.D
 Acq On : 18 Jun 2018 13:55
 Operator : MD\SY
 Sample : J3523-05ME
 Misc : 4.00g/5mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 AREA1(D)ME

Quant Time: Jun 19 02:15:55 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:02:01 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	1193506	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1803668	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1624193	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	660747	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	753230	47.58	ug/l	0.00
Spiked Amount			Recovery	=	95.16%	
35) Dibromofluoromethane	7.59	113	659790	45.21	ug/l	0.00
Spiked Amount			Recovery	=	90.42%	
50) Toluene-d8	10.09	98	2544368	45.79	ug/l	0.00
Spiked Amount			Recovery	=	91.58%	
62) 4-Bromofluorobenzene	12.40	95	822013	44.74	ug/l	0.00
Spiked Amount			Recovery	=	89.48%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.83	43	8212	2.137	ug/l	90
17) Carbon Disulfide	4.03	76	18989	0.443	ug/l	98
18) Methyl Acetate	4.33	43	137440	11.728	ug/l	95
19) Methyl tert-butyl Ether	5.05	73	76883	2.204	ug/l	94
20) Methylene Chloride	4.54	84	9623	0.599	ug/l #	78
39) Methylcyclohexane	9.08	83	11734	0.529	ug/l	93
40) Benzene	8.04	78	54842	0.865	ug/l	99
52) Toluene	10.16	92	31149	0.837	ug/l	95
68) m/p-Xylenes	11.62	106	20396	0.783	ug/l	90
69) o-Xylene	11.95	106	6095	0.243	ug/l	98
80) 1,3,5-Trimethylbenzene	12.73	105	13428	0.294	ug/l	91
84) 1,2,4-Trimethylbenzene	13.04	105	24552	0.542	ug/l	99
95) Naphthalene	15.14	128	119618	8.495	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN061818\
Data File : VN049340.D
Acq On : 18 Jun 2018 13:55
Operator : MD\SY
Sample : J3523-05ME
Misc : 4.00µL/5mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
AREA1(D)ME

Quant Time: Jun 19 02:15:55 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:02:01 2018
Response via : Initial Calibration

