

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061818\
 Data File : VN049346.D
 Acq On : 18 Jun 2018 16:30
 Operator : MD\SY
 Sample : PB110313ZHE#11
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB110313ZHE#11

Quant Time: Jun 19 03:10:38 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:02:01 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1335864	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	2074033	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1682849	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	469822	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	874372	49.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.70%	
35) Dibromofluoromethane	7.59	113	759327	45.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.50%	
50) Toluene-d8	10.09	98	2828316	44.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.54%	
62) 4-Bromofluorobenzene	12.41	95	717057	33.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	67.88%	
Target Compounds						
16) Acetone	3.82	43	60159	13.984	ug/l #	89
18) Methyl Acetate	4.33	43	122007	9.301	ug/l	96
20) Methylene Chloride	4.55	84	69993	3.896	ug/l	87
95) Naphthalene	15.14	128	30955	4.443	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN061818\
Data File : VN049346.D
Acq On : 18 Jun 2018 16:30
Operator : MD\SY
Sample : PB110313ZHE#11
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB110313ZHE#11

Quant Time: Jun 19 03:10:38 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:02:01 2018
Response via : Initial Calibration

