

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061818\
 Data File : VN049348.D
 Acq On : 18 Jun 2018 17:21
 Operator : MD\SY
 Sample : PB110313ZHE#13
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB110313ZHE#13

Quant Time: Jun 19 03:19:23 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:02:01 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1297396	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	2055103	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1672355	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	543227	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	848820	49.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.66%	
35) Dibromofluoromethane	7.59	113	756021	45.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.94%	
50) Toluene-d8	10.09	98	2770765	43.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.54%	
62) 4-Bromofluorobenzene	12.40	95	744141	35.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	71.10%	
Target Compounds						
16) Acetone	3.82	43	55619	13.312	ug/l #	88
18) Methyl Acetate	4.33	43	148562	11.662	ug/l	97
20) Methylene Chloride	4.55	84	67604	3.874	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN061818\
Data File : VN049348.D
Acq On : 18 Jun 2018 17:21
Operator : MD\SY
Sample : PB110313ZHE#13
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB110313ZHE#13

Quant Time: Jun 19 03:19:23 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:02:01 2018
Response via : Initial Calibration

