

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061818\
 Data File : VN049353.D
 Acq On : 18 Jun 2018 19:31
 Operator : MD\SY
 Sample : PB110313ZHE#18
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB110313ZHE#18

Quant Time: Jun 19 03:34:08 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:02:01 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1262451	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	2030283	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1651168	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	528396	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	844307	50.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.84%	
35) Dibromofluoromethane	7.59	113	757054	46.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.18%	
50) Toluene-d8	10.09	98	2740258	43.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.62%	
62) 4-Bromofluorobenzene	12.41	95	729109	35.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	70.52%	
Target Compounds						
16) Acetone	3.82	43	56650	13.934	ug/l	92
18) Methyl Acetate	4.33	43	153333	12.369	ug/l	95
20) Methylene Chloride	4.55	84	66987	3.945	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN061818\
Data File : VN049353.D
Acq On : 18 Jun 2018 19:31
Operator : MD\SY
Sample : PB110313ZHE#18
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB110313ZHE#18

Quant Time: Jun 19 03:34:08 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:02:01 2018
Response via : Initial Calibration

