

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061818\
 Data File : VN049356.D
 Acq On : 18 Jun 2018 20:48
 Operator : MD\SY
 Sample : PB110313ZHE#09
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB110313ZHE#09

Quant Time: Jun 19 03:40:05 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:02:01 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1204379	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1935614	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1549243	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	468978	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	811813	50.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.64%	
35) Dibromofluoromethane	7.59	113	716207	45.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.46%	
50) Toluene-d8	10.09	98	2580598	43.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.56%	
62) 4-Bromofluorobenzene	12.41	95	663175	33.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	67.28%	
Target Compounds						
16) Acetone	3.83	43	56619	14.598	ug/l	93
18) Methyl Acetate	4.33	43	159617	13.497	ug/l	94
20) Methylene Chloride	4.55	84	67357	4.158	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN061818\
Data File : VN049356.D
Acq On : 18 Jun 2018 20:48
Operator : MD\SY
Sample : PB110313ZHE#09
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB110313ZHE#09

Quant Time: Jun 19 03:40:05 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:02:01 2018
Response via : Initial Calibration

