

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061818\
 Data File : VN049358.D
 Acq On : 18 Jun 2018 21:40
 Operator : MD\SY
 Sample : J3523-07ME
 Misc : 4.01g/5mL/100uL/5mL/MSVOA_N/MEOH
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 AREA2(D)ME

Quant Time: Jun 19 13:01:31 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:02:01 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	1158507	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1864111	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1574395	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	656112	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	772611	50.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.56%	
35) Dibromofluoromethane	7.59	113	679466	45.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.10%	
50) Toluene-d8	10.09	98	2531215	44.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.16%	
62) 4-Bromofluorobenzene	12.40	95	775581	40.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	81.70%	
Target Compounds						
40) Benzene	8.04	78	228219	3.484	ug/l	99
52) Toluene	10.16	92	46147	1.200	ug/l	98
95) Naphthalene	15.14	128	124889	8.818	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN061818\
Data File : VN049358.D
Acq On : 18 Jun 2018 21:40
Operator : MD\SY
Sample : J3523-07ME
Misc : 4.01µL/5mL/100µL/5mL/MSVOA_N/MEOH
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
AREA2(D)ME

Quant Time: Jun 19 13:01:31 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:02:01 2018
Response via : Initial Calibration

