

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061818\
 Data File : VN049371.D
 Acq On : 19 Jun 2018 4:06
 Operator : MD\SY
 Sample : J3567-12
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-11-S

Quant Time: Jun 19 08:08:19 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:02:01 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1282840	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	2048058	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1682576	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	532583	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	853443	50.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.32%	
35) Dibromofluoromethane	7.59	113	752428	45.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.82%	
50) Toluene-d8	10.09	98	2784227	44.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.26%	
62) 4-Bromofluorobenzene	12.41	95	739640	35.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	70.90%	
Target Compounds						
16) Acetone	3.82	43	33319	8.065	ug/l	96
18) Methyl Acetate	4.33	43	126044	10.006	ug/l	98
20) Methylene Chloride	4.55	84	47790	2.770	ug/l	88
95) Naphthalene	15.14	128	80673	7.464	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN061818\
Data File : VN049371.D
Acq On : 19 Jun 2018 4:06
Operator : MD\SY
Sample : J3567-12
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-11-S

Quant Time: Jun 19 08:08:19 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:02:01 2018
Response via : Initial Calibration

