

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061819\
 Data File : VN056356.D
 Acq On : 18 Jun 2019 11:21
 Operator : JC/SP
 Sample : K3163-12
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SU-03-061719-B

Quant Time: Jun 19 02:25:33 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061719W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 18 04:23:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	334741	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	567750	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	590667	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	196379	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	195817	52.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.06%	
35) Dibromofluoromethane	7.59	113	170976	46.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.92%	
50) Toluene-d8	10.09	98	730044	52.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.98%	
62) 4-Bromofluorobenzene	12.40	95	246897	53.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.14%	
Target Compounds						
16) Acetone	3.83	43	10704	5.147	ug/l	96
20) Methylene Chloride	4.54	84	130743	35.601	ug/l	99
43) Isopropyl Acetate	8.17	43	13087	1.690	ug/l	95
95) Naphthalene	15.13	128	110813	15.845	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN061819\
Data File : VN056356.D
Acq On : 18 Jun 2019 11:21
Operator : JC/SP
Sample : K3163-12
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SU-03-061719-B

Quant Time: Jun 19 02:25:33 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061719W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 18 04:23:59 2019
Response via : Initial Calibration

