

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061819\
 Data File : VN056361.D
 Acq On : 18 Jun 2019 13:20
 Operator : JC/SP
 Sample : K3157-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 CL-01-061719-B

Quant Time: Jun 19 03:15:52 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061719W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 18 04:23:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	332404	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	568889	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	583125	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	184803	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	195533	52.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.64%	
35) Dibromofluoromethane	7.59	113	169978	46.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.18%	
50) Toluene-d8	10.09	98	721055	51.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.48%	
62) 4-Bromofluorobenzene	12.40	95	243852	52.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.60%	
Target Compounds						
16) Acetone	3.82	43	14879	7.204	ug/l	100
20) Methylene Chloride	4.54	84	26186	7.181	ug/l	98
27) cis-1,2-Dichloroethene	6.83	96	14694	3.866	ug/l	90
43) Isopropyl Acetate	8.17	43	11994	1.545	ug/l	96
44) Trichloroethene	8.83	130	22113	5.236	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN061819\
Data File : VN056361.D
Acq On : 18 Jun 2019 13:20
Operator : JC/SP
Sample : K3157-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
CL-01-061719-B

Quant Time: Jun 19 03:15:52 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061719W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 18 04:23:59 2019
Response via : Initial Calibration

