

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061819\
 Data File : VN056363.D
 Acq On : 18 Jun 2019 14:08
 Operator : JC/SP
 Sample : K3157-10
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 CL-01-061719-E

Quant Time: Jun 19 03:21:34 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061719W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 18 04:23:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	346325	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	590914	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	606137	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	190052	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	203595	52.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.56%	
35) Dibromofluoromethane	7.59	113	175458	46.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.60%	
50) Toluene-d8	10.09	98	752419	51.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.96%	
62) 4-Bromofluorobenzene	12.40	95	250135	52.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.28%	
Target Compounds						
20) Methylene Chloride	4.55	84	11471	3.019	ug/l	98
43) Isopropyl Acetate	8.17	43	12709	1.577	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN061819\
Data File : VN056363.D
Acq On : 18 Jun 2019 14:08
Operator : JC/SP
Sample : K3157-10
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
CL-01-061719-E

Quant Time: Jun 19 03:21:34 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061719W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 18 04:23:59 2019
Response via : Initial Calibration

