


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN061819\
Data File : VN056349.D
Acq On : 18 Jun 2019 8:21
Operator : JC/SP
Sample : BFB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061719W.M
Title : SW846 8260
Last Update : Tue Jun 18 04:23:59 2019



AutoFind: Scans 3304, 3305, 3306; Background Corrected with Scan 3291

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.7	7941	PASS
75	95	30	60	46.3	20821	PASS
95	95	100	100	100.0	44936	PASS
96	95	5	9	6.2	2772	PASS
173	174	0.00	2	1.3	528	PASS
174	95	50	100	91.2	40992	PASS
175	174	5	9	6.8	2782	PASS
176	174	95	101	95.1	38992	PASS
177	176	5	9	6.4	2483	PASS