

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061820\
 Data File : VN061924.D
 Acq On : 19 Jun 2020 2:05
 Operator : JC/MD
 Sample : PB129596ZHE#01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB129596ZHE#01

Quant Time: Jun 19 05:33:32 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 17 01:19:44 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	152277	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	281940	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	281094	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	103243	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	7.99	65	102990	50.43	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.86%	
35) Dibromofluoromethane	7.55	113	89132	49.42	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.84%	
50) Toluene-d8	10.06	98	364363	51.38	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.76%	
62) 4-Bromofluorobenzene	12.38	95	119608	49.17	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.34%	

Target Compounds


						Qvalue
16) Acetone	3.78	43	10323	17.112	ug/l	95
18) Methyl Acetate	4.29	43	1927	1.110	ug/l #	81
20) Methylene Chloride	4.50	84	4365	2.221	ug/l	96
25) 2-Butanone	6.80	43	1513	1.719	ug/l #	83
29) Tetrahydrofuran	7.17	42	616	1.063	ug/l #	40
31) Cyclohexane	7.62	56	3218	1.196	ug/l #	10
36) 1,1-Dichloropropene	7.63	75	11881	4.620	ug/l #	48
37) Ethyl Acetate	6.80	43	1513	0.753	ug/l #	65
38) Carbon Tetrachloride	7.63	117	15035	5.904	ug/l #	17
43) Isopropyl Acetate	8.13	43	35351	10.446	ug/l #	89
45) 1,2-Dichloropropane	9.24	63	796	0.416	ug/l #	43
48) Methyl methacrylate	9.14	41	1907	1.239	ug/l #	5
51) 4-Methyl-2-Pentanone	9.95	43	5213	2.607	ug/l	89
54) cis-1,3-Dichloropropene	9.88	75	18193	5.716	ug/l #	69
56) Ethyl methacrylate	10.52	69	797	3.158	ug/l #	1
57) 1,3-Dichloropropane	10.74	76	3502	1.074	ug/l #	42
59) 2-Hexanone	10.74	43	40460	37.372	ug/l #	48
71) Bromoform	12.37	173	606	0.402	ug/l #	1
74) N-amyl acetate	12.00	43	2076	3.798	ug/l #	38
76) 1,2,3-Trichloropropane	12.38	75	59976	29.590	ug/l #	100
80) 1,3,5-Trimethylbenzene	12.71	105	2247	0.370	ug/l	96
81) trans-1,4-Dichloro-2-buten	12.38	75	59976	80.704	ug/l #	14
95) Naphthalene	15.12	128	62	3.494	ug/l #	69

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN061820\
Data File : VN061924.D
Acq On : 19 Jun 2020 2:05
Operator : JC/MD
Sample : PB129596ZHE#01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB129596ZHE#01

Quant Time: Jun 19 05:33:32 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061620W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 17 01:19:44 2020
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.63 min Scan# 1820

Delta R.T. 0.00 min

Lab File: VN061924.D

Acq: 19 Jun 2020 2:05

Instrument :

MSVOA_N

ClientSampleId :


PB129596ZHE#01

Tot Ion:168 Resp: 152277

Ion Ratio Lower Upper




168 100

99 57.1 44.6 67.0



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VN



#16

Acetone

Concen: 17.112 ug/l

RT: 3.78 min Scan# 624

Delta R.T. 0.01 min

Lab File: VN061924.D


Acq: 19 Jun 2020 2:05

Tgt Ion: 43 Resp: 10323

Ion Ratio Lower Upper



43 100


58 37.8 27.8 41.8



Abundance Ion 43.00 (42.70 to 43.70): VN

Ion 58.00 (57.70 to 58.70): VN









#18
Methyl Acetate
Concen: 1.110 ug/l
RT: 4.29 min Scan# 783
Delta R.T. 0.02 min
Lab File: VN061924.D
Acq: 19 Jun 2020 2:05




Instrument :
MSVOA_N
ClientSampleId :
PB129596ZHE#01


Tot Ion: 43 Resp: 1927
Ion Ratio Lower Upper
43 100
74 18.5 23.0 34.4#



#20
Methylene Chloride
Concen: 2.221 ug/l
RT: 4.50 min Scan# 848
Delta R.T. 0.00 min
Lab File: VN061924.D
Acq: 19 Jun 2020 2:05

Tgt Ion: 84 Resp: 4365
Ion Ratio Lower Upper
84 100
49 103.4 87.0 130.4
51 31.6 27.8 41.6
86 61.4 50.2 75.2









#25
2-Butanone
Concen: 1.719 ug/l
RT: 6.80 min Scan# 1563
Delta R.T. 0.02 min
Lab File: VN061924.D
Acq: 19 Jun 2020 2:05




Instrument :
MSVOA_N
ClientSampleId :
PB129596ZHE#01


Tot Ion: 43 Resp: 1513
Ion Ratio Lower Upper
43 100
72 21.7 24.6 37.0#



#29
Tetrahydrofuran
Concen: 1.063 ug/l
RT: 7.17 min Scan# 1678
Delta R.T. 0.00 min
Lab File: VN061924.D
Acq: 19 Jun 2020 2:05

Tgt Ion: 42 Resp: 616
Ion Ratio Lower Upper
42 100
72 15.6 44.0 66.0#
71 4.7 40.0 60.0#





#31

Cyclohexane

Concen: 1.196 ug/l

RT: 7.62 min Scan# 1819

Delta R.T. 0.01 min

Lab File: VN061924.D

Acq: 19 Jun 2020 2:05

Instrument :

MSVOA_N

ClientSampleId :

PB129596ZHE#01


Tot Ion: 56 Resp: 3218

Ion Ratio Lower Upper

56 100

69 168.9 27.7 41.5#




84 138.2 79.1 118.7#



Abundance Ion 56.00 (55.70 to 56.70): VN061834.D (-1795) (-)

Ion 69.00 (68.70 to 69.70): VN061834.D (-1795) (-)

Ion 84.00 (83.70 to 84.70): VN061834.D (-1795) (-)



#33

1,2-Dichloroethane-d4

Concen: 50.429 ug/l

RT: 7.99 min Scan# 1932

Delta R.T. -0.00 min

Lab File: VN061924.D


Acq: 19 Jun 2020 2:05

Tgt Ion: 65 Resp: 102990

Ion Ratio Lower Upper



65 100


67 51.6 0.0 106.2



Abundance Ion 64.90 (64.60 to 65.60): VN061834.D (-1915) (-)

Ion 66.90 (66.60 to 67.60): VN061834.D (-1915) (-)





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.55 min Scan# 2108

Delta R.T. 0.00 min

Lab File: VN061924.D

Acq: 19 Jun 2020 2:05

Instrument :

MSVOA_N

ClientSampleId :

PB129596ZHE#01


Tot Ion:114 Resp: 281940

Ion Ratio Lower Upper

114 100

63 18.6 0.0 37.6




88 15.4 0.0 31.0



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VN

Ion 87.90 (87.60 to 88.60): VN



#35

Dibromofluoromethane

Concen: 49.416 ug/l

RT: 7.55 min Scan# 1796

Delta R.T. 0.00 min

Lab File: VN061924.D

Acq: 19 Jun 2020 2:05


Tgt Ion:113 Resp: 89132

Ion Ratio Lower Upper

113 100

111 103.4 83.7 125.5



192 17.7 13.9 20.9




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V





#36

1,1-Dichloropropene

Concen: 4.620 ug/l

RT: 7.63 min Scan# 1820

Delta R.T. -0.13 min

Lab File: VN061924.D

Acq: 19 Jun 2020 2:05

Instrument :

MSVOA_N

ClientSampleId :

PB129596ZHE#01


Tot Ion: 75 Resp: 11881

Ion Ratio Lower Upper

75 100

110 7.1 18.2 54.6#




77 0.0 24.2 36.4#



Abundance Ion 74.90 (74.60 to 75.60): VN0

Ion 109.90 (109.60 to 110.60): VN0

Ion 76.90 (76.60 to 77.60): VN0



#37

Ethyl Acetate

Concen: 0.753 ug/l

RT: 6.80 min Scan# 1563

Delta R.T. -0.08 min

Lab File: VN061924.D

Acq: 19 Jun 2020 2:05


Tgt Ion: 43 Resp: 1513

Ion Ratio Lower Upper

43 100

61 0.0 13.0 19.4#



70 0.0 9.4 14.2#




Abundance Ion 43.00 (42.70 to 43.70): VN0

Ion 60.90 (60.60 to 61.60): VN0

Ion 70.00 (69.70 to 70.70): VN0





#38

Carbon Tetrachloride

Concen: 5.904 ug/l

RT: 7.63 min Scan# 1821

Delta R.T. -0.11 min

Lab File: VN061924.D

Acq: 19 Jun 2020 2:05

Instrument :

MSVOA_N

ClientSampleId :

PB129596ZHE#01


Tot Ion:117 Resp: 15035

Ion Ratio Lower Upper

117 100

119 5.2 75.8 113.6#




121 0.0 23.8 35.6#



Abundance Ion 116.80 (116.50 to 117.50): V

Ion 118.80 (118.50 to 119.50): V

Ion 120.80 (120.50 to 121.50): V



#43

Isopropyl Acetate

Concen: 10.446 ug/l

RT: 8.13 min Scan# 1976

Delta R.T. 0.00 min

Lab File: VN061924.D

Acq: 19 Jun 2020 2:05


Tgt Ion: 43 Resp: 35351

Ion Ratio Lower Upper

43 100

61 20.8 24.0 36.0#



87 15.3 12.5 18.7




Abundance Ion 43.00 (42.70 to 43.70): VN0

Ion 61.00 (60.70 to 61.70): VN0

Ion 87.00 (86.70 to 87.70): VN0









#45
 1,2-Dichloropropane
 Concen: 0.416 ug/l
 RT: 9.24 min Scan# 2321
 Delta R.T. 0.15 min
 Lab File: VN061924.D
 Acq: 19 Jun 2020 2:05




Instrument :
 MSVOA_N
 ClientSampleId :
 PB129596ZHE#01


Tot Ion: 63 Resp: 796
 Ion Ratio Lower Upper
 63 100
 65 0.0 25.1 37.7#



#48
 Methyl methacrylate
 Concen: 1.239 ug/l
 RT: 9.14 min Scan# 2292
 Delta R.T. -0.02 min
 Lab File: VN061924.D
 Acq: 19 Jun 2020 2:05

Tgt Ion: 41 Resp: 1907
 Ion Ratio Lower Upper
 41 100
 69 0.0 83.0 124.4#
 39 0.0 55.1 82.7#





#50

Toluene-d8

Concen: 51.383 ug/l

RT: 10.06 min Scan# 2577

Delta R.T. 0.00 min

Lab File: VN061924.D

Acq: 19 Jun 2020 2:05

Instrument :

MSVOA_N

ClientSampleId :


PB129596ZHE#01

Tot Ion: 98 Resp: 364363




Ion Ratio Lower Upper

98 100

100 64.9 52.7 79.1



Abundance Ion 98.00 (97.70 to 98.70): VN061834.D (-2561) (-)



#51

4-Methyl-2-Pentanone

Concen: 2.607 ug/l

RT: 9.95 min Scan# 2544

Delta R.T. 0.00 min

Lab File: VN061924.D


Acq: 19 Jun 2020 2:05

Tgt Ion: 43 Resp: 5213



Ion Ratio Lower Upper


43 100

58 36.4 34.6 52.0



Abundance Ion 43.00 (42.70 to 43.70): VN061834.D (-2527) (-)






#54
 cis-1,3-Dichloropropene
 Concen: 5.716 ug/l
 RT: 9.88 min Scan# 2520
 Delta R.T. 0.07 min
 Lab File: VN061924.D
 Acq: 19 Jun 2020 2:05

Instrument :
 MSVOA_N
 ClientSampleId :
 PB129596ZHE#01

Tot Ion	Ratio	Lower	Upper
75	100		
77	0.0	24.8	37.2#
39	35.3	35.4	53.0#






Abundance

Ion 74.90 (74.60 to 75.60): VN0


Ion 76.90 (76.60 to 77.60): VN0

Ion 39.00 (38.70 to 39.70): VN0



#56
 Ethyl methacrylate
 Concen: 3.158 ug/l
 RT: 10.52 min Scan# 2719
 Delta R.T. 0.12 min
 Lab File: VN061924.D
 Acq: 19 Jun 2020 2:05

Tgt Ion	Ratio	Lower	Upper
69	100		
41	4674.7	44.9	67.3#
39	2253.1	30.0	45.0#






Abundance

Ion 69.00 (68.70 to 69.70): VN0

Ion 41.00 (40.70 to 41.70): VN0

Ion 39.00 (38.70 to 39.70): VN0









#57
 1,3-Dichloropropane
 Concen: 1.074 ug/l
 RT: 10.74 min Scan# 2788
 Delta R.T. 0.06 min
 Lab File: VN061924.D
 Acq: 19 Jun 2020 2:05




Instrument :
 MSVOA_N
 ClientSampleId :
 PB129596ZHE#01


Tot Ion: 76 Resp: 3502
 Ion Ratio Lower Upper
 76 100
 78 0.0 25.8 38.8#



#59
 2-Hexanone
 Concen: 37.372 ug/l
 RT: 10.74 min Scan# 2788
 Delta R.T. 0.02 min
 Lab File: VN061924.D
 Acq: 19 Jun 2020 2:05

Tgt Ion: 43 Resp: 40460
 Ion Ratio Lower Upper
 43 100
 58 20.7 30.1 90.5#









#62
4-Bromofluorobenzene
Concen: 49.174 ug/l
RT: 12.38 min Scan# 3298
Delta R.T. 0.00 min
Lab File: VN061924.D
Acq: 19 Jun 2020 2:05

Instrument :
MSVOA_N
ClientSampleId :
PB129596ZHE#01

Tot Ion: 95 Resp: 119608
Ion Ratio Lower Upper
95 100
174 79.4 0.0 158.6
176 74.3 0.0 152.0




Abundance Ion 94.90 (94.60 to 95.60): VN061834.D (-3284) (-)
Ion 173.80 (173.50 to 174.50): VN061834.D (-3284) (-)
Ion 175.80 (175.50 to 176.50): VN061834.D (-3284) (-)






#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.38 min Scan# 2988
Delta R.T. 0.00 min
Lab File: VN061924.D
Acq: 19 Jun 2020 2:05

Tgt Ion: 117 Resp: 281094
Ion Ratio Lower Upper
117 100
82 53.9 42.6 64.0
119 32.1 24.7 37.1



Abundance Ion 116.90 (116.60 to 117.60): VN061834.D (-2973) (-)
Ion 82.00 (81.70 to 82.70): VN061834.D (-2973) (-)
Ion 118.90 (118.60 to 119.60): VN061834.D (-2973) (-)





#71

Bromoform

Concen: 0.402 ug/l

RT: 12.37 min Scan# 3297

Delta R.T. 0.27 min

Lab File: VN061924.D

Acq: 19 Jun 2020 2:05

Instrument :

MSVOA_N

ClientSampleId :

PB129596ZHE#01


Tot Ion:173 Resp: 606

Ion Ratio Lower Upper

173 100

175 1028.7 23.9 71.7#




254 0.0 0.0 0.0



Abundance Ion 172.70 (172.40 to 173.40): V

Ion 174.70 (174.40 to 175.40): V

Ion 254.40 (254.10 to 255.10): V



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.32 min Scan# 3591

Delta R.T. 0.00 min

Lab File: VN061924.D

Acq: 19 Jun 2020 2:05


Tgt Ion:152 Resp: 103243

Ion Ratio Lower Upper

152 100

115 59.4 29.1 87.3



150 158.1 0.0 347.6




Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V





#74

N-amvl acetate

Concen: 3.798 ug/l

RT: 12.00 min Scan# 3182

Delta R.T. -0.04 min

Lab File: VN061924.D

Acq: 19 Jun 2020 2:05

Instrument :

MSVOA_N

ClientSampleId :

PB129596ZHE#01

Tot Ion: 43 Resp: 2076



Ion Ratio Lower Upper

43 100

70 0.0 40.6 60.8#

55 0.0 21.5 32.3#

61 0.0 21.6 32.4#





Abundance Ion 43.00 (42.70 to 43.70): VN0

Ion 70.00 (69.70 to 70.70): VN0

Ion 55.00 (54.70 to 55.70): VN0

Ion 61.00 (60.70 to 61.70): VN0



#76

1,2,3-Trichloropropane

Concen: 29.590 ug/l

RT: 12.38 min Scan# 3298

Delta R.T. -0.15 min

Lab File: VN061924.D



Acq: 19 Jun 2020 2:05

Tgt Ion: 75 Resp: 59976

Ion Ratio Lower Upper


75 100


77 1.2 0.0 0.0#



Abundance Ion 74.90 (74.60 to 75.60): VN0

Ion 77.00 (76.70 to 77.70): VN0









#80
 1,3,5-Trimethylbenzene
 Concen: 0.370 ug/l
 RT: 12.71 min Scan# 3401
 Delta R.T. 0.00 min
 Lab File: VN061924.D
 Acq: 19 Jun 2020 2:05




Instrument :
 MSVOA_N
 ClientSampleId :
 PB129596ZHE#01


Tot Ion:105 Resp: 2247
 Ion Ratio Lower Upper
 105 100
 120 47.0 24.9 74.7



#81
 trans-1,4-Dichloro-2-butene
 Concen: 80.704 ug/l
 RT: 12.38 min Scan# 3298
 Delta R.T. 0.10 min
 Lab File: VN061924.D
 Acq: 19 Jun 2020 2:05

Tgt Ion: 75 Resp: 59976
 Ion Ratio Lower Upper
 75 100
 53 0.0 70.4 105.6#
 89 0.0 36.0 54.0#





#95

Naphthalene

Concen: 3.494 ug/l

RT: 15.12 min Scan# 4150

Delta R.T. 0.01 min

Lab File: VN061924.D

Acq: 19 Jun 2020 2:05

Instrument :

MSVOA_N

ClientSampleId :

PB129596ZHE#01


Tot Ion:128 Resp: 62

Ion Ratio Lower Upper

128 100

127 0.0 10.3 15.5#

129 0.0 8.6 13.0#



Abundance Ion 128.00 (127.70 to 128.70): V

Ion 127.00 (126.70 to 127.70): V

Ion 129.00 (128.70 to 129.70): V

