

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061820\
 Data File : VN061898.D
 Acq On : 18 Jun 2020 14:26
 Operator : JC/MD
 Sample : L3011-16
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE104D1-20200612

Quant Time: Jun 19 05:57:14 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 17 01:19:44 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	130429	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	237525	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	227565	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	83526	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	88255	50.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.90%	
35) Dibromofluoromethane	7.55	113	78602	51.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.46%	
50) Toluene-d8	10.06	98	297228	49.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.50%	
62) 4-Bromofluorobenzene	12.38	95	94415	46.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.14%	
Target Compounds						
9) 1,1,2-Trichlorotrifluoroet	3.72	101	1616	1.222	ug/l #	67
16) Acetone	3.78	43	2459	4.759	ug/l	90
44) Trichloroethene	8.80	130	87628	50.735	ug/l	100
64) Tetrachloroethene	10.60	164	4456	3.072	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061820\
Data File : VN061898.D
Acq On : 18 Jun 2020 14:26
Operator : JC/MD
Sample : L3011-16
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE104D1-20200612

Quant Time: Jun 19 05:57:14 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061620W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 17 01:19:44 2020
Response via : Initial Calibration

