

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061820\
 Data File : VN061902.D
 Acq On : 18 Jun 2020 16:03
 Operator : JC/MD
 Sample : L3011-20
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE-DUP04-20200612

Quant Time: Jun 19 06:05:48 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 17 01:19:44 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	135845	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	248191	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	237986	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	89987	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	91651	50.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.62%	
35) Dibromofluoromethane	7.55	113	80136	50.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.94%	
50) Toluene-d8	10.06	98	307966	49.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.68%	
62) 4-Bromofluorobenzene	12.38	95	101032	47.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.36%	
Target Compounds						
16) Acetone	3.78	43	3508	6.519	ug/l #	85
44) Trichloroethene	8.80	130	15374	8.519	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN061820\
Data File : VN061902.D
Acq On : 18 Jun 2020 16:03
Operator : JC/MD
Sample : L3011-20
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE-DUP04-20200612

Quant Time: Jun 19 06:05:48 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061620W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 17 01:19:44 2020
Response via : Initial Calibration

