

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061820\
 Data File : VN061903.D
 Acq On : 18 Jun 2020 16:27
 Operator : JC/MD
 Sample : L3026-07
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 S49-TB-2020-3

Quant Time: Jun 19 06:08:20 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 17 01:19:44 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	133571	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	245820	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	238544	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	92861	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	7.99	65	92508	51.64	ug/l	0.00
Spiked Amount			Recovery	=	103.28%	
35) Dibromofluoromethane	7.55	113	80385	51.12	ug/l	0.00
Spiked Amount			Recovery	=	102.24%	
50) Toluene-d8	10.06	98	306814	49.63	ug/l	0.00
Spiked Amount			Recovery	=	99.26%	
62) 4-Bromofluorobenzene	12.38	95	101878	48.04	ug/l	0.00
Spiked Amount			Recovery	=	96.08%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN061820\
Data File : VN061903.D
Acq On : 18 Jun 2020 16:27
Operator : JC/MD
Sample : L3026-07
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
S49-TB-2020-3

Quant Time: Jun 19 06:08:20 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061620W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 17 01:19:44 2020
Response via : Initial Calibration

