

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061820\
 Data File : VN061925.D
 Acq On : 19 Jun 2020 2:29
 Operator : JC/MD
 Sample : PB129596ZHE#02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB129596ZHE#02

Quant Time: Jun 19 07:19:56 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 17 01:19:44 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	145998	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	274269	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	271346	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	103681	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	97865	49.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.96%	
35) Dibromofluoromethane	7.55	113	87304	49.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.52%	
50) Toluene-d8	10.06	98	352172	51.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.10%	
62) 4-Bromofluorobenzene	12.38	95	114713	48.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.96%	
Target Compounds						
16) Acetone	3.78	43	10071	17.413	ug/l	96
18) Methyl Acetate	4.28	43	2344	1.408	ug/l #	68
20) Methylene Chloride	4.51	84	4470	2.372	ug/l	88
43) Isopropyl Acetate	8.13	43	34085	10.353	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN061820\
Data File : VN061925.D
Acq On : 19 Jun 2020 2:29
Operator : JC/MD
Sample : PB129596ZHE#02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB129596ZHE#02

Quant Time: Jun 19 07:19:56 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061620W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 17 01:19:44 2020
Response via : Initial Calibration

