

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061820\
 Data File : VN061934.D
 Acq On : 19 Jun 2020 6:06
 Operator : JC/MD
 Sample : PB129596ZHE#11
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB129596ZHE#11

Quant Time: Jun 19 07:57:17 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 17 01:19:44 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	139490	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	266835	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	267862	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	100747	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	95118	50.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.68%	
35) Dibromofluoromethane	7.55	113	84632	49.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.16%	
50) Toluene-d8	10.06	98	340174	50.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.38%	
62) 4-Bromofluorobenzene	12.38	95	112987	49.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.16%	
Target Compounds						
16) Acetone	3.78	43	10837	19.611	ug/l	92
18) Methyl Acetate	4.28	43	4241	2.666	ug/l	91
20) Methylene Chloride	4.51	84	4543	2.523	ug/l	87
43) Isopropyl Acetate	8.13	43	33471	10.450	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN061820\
Data File : VN061934.D
Acq On : 19 Jun 2020 6:06
Operator : JC/MD
Sample : PB129596ZHE#11
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 50 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB129596ZHE#11

Quant Time: Jun 19 07:57:17 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061620W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 17 01:19:44 2020
Response via : Initial Calibration

