

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061821\
 Data File : VN067392.D
 Acq On : 18 Jun 2021 11:07
 Operator : JC/MD
 Sample : VN0618WBL01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0618WBL01

Quant Time: Jun 19 00:36:12 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061421W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 15 05:19:42 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	316594	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	640939	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	579207	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	201779	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	303113	48.533	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	97.060%
35) Dibromofluoromethane	8.024	113	195131	45.776	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	91.560%
50) Toluene-d8	10.443	98	848874	51.956	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	103.920%
62) 4-Bromofluorobenzene	12.734	95	297774	44.163	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	88.320%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061821\
Data File : VN067392.D
Acq On : 18 Jun 2021 11:07
Operator : JC/MD
Sample : VN0618WBL01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0618WBL01

Quant Time: Jun 19 00:36:12 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061421W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 15 05:19:42 2021
Response via : Initial Calibration

