

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061821\
 Data File : VN067399.D
 Acq On : 18 Jun 2021 14:49
 Operator : JC/MD
 Sample : PB137182ZHE#01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB137182ZHE#01

Quant Time: Jun 19 00:39:46 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061421W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 15 05:19:42 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	318023	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	638341	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	573382	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	202118	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	304683	48.565	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	97.120%
35) Dibromofluoromethane	8.027	113	197106	46.427	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	92.860%
50) Toluene-d8	10.443	98	841484	51.713	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	103.420%
62) 4-Bromofluorobenzene	12.734	95	297354	44.280	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	88.560%
Target Compounds						
					Qvalue	
16) Acetone	4.301	43	29934	12.433	ug/l	95
20) Methylene Chloride	5.103	84	20301	4.053	ug/l	93
43) Isopropyl Acetate	8.566	43	58395	4.291	ug/l #	93
95) Naphthalene	15.515	128	8157	3.237	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061821\
Data File : VN067399.D
Acq On : 18 Jun 2021 14:49
Operator : JC/MD
Sample : PB137182ZHE#01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB137182ZHE#01

Quant Time: Jun 19 00:39:46 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061421W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 15 05:19:42 2021
Response via : Initial Calibration

