

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061821\
 Data File : VN067402.D
 Acq On : 18 Jun 2021 16:02
 Operator : JC/MD
 Sample : M2762-02 50X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC_AQUEOUS_DBD

Manual Integrations
 APPROVED

MMDadoda
 6/21/2021 3:47:26 PM

Quant Time: Jun 19 00:40:40 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061421W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 15 05:19:42 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	294418	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	599454	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	527950	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	186031	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	283265	48.771	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	97.540%		
35) Dibromofluoromethane	8.026	113	182995	45.900	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	91.800%		
50) Toluene-d8	10.446	98	777815	50.901	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	101.800%		
62) 4-Bromofluorobenzene	12.733	95	274993	43.607	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	87.220%		
Target Compounds						Qvalue
16) Acetone	4.293	43	12466m	5.593	ug/l	
18) Methyl Acetate	4.870	43	11489m	2.126	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061821\
Data File : VN067402.D
Acq On : 18 Jun 2021 16:02
Operator : JC/MD
Sample : M2762-02 50X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC_AQUEOUS_DBD

Manual Integrations
APPROVED

MMDadoda
6/21/2021 3:47:26 PM

Quant Time: Jun 19 00:40:40 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061421W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 15 05:19:42 2021
Response via : Initial Calibration

