

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061825\
 Data File : VN087075.D
 Acq On : 18 Jun 2025 11:01
 Operator : JC\MD
 Sample : VN0618MBL01
 Misc : 5.00g/10mL/100uL/5.0mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0618MBL01

Quant Time: Jun 19 01:23:57 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 02:12:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	179681	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	357376	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	315794	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	149459	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	127106	52.835	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	105.660%
35) Dibromofluoromethane	8.177	113	109311	51.613	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.220%
50) Toluene-d8	10.571	98	439977	52.477	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	104.960%
62) 4-Bromofluorobenzene	12.847	95	150478	48.308	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	96.620%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061825\
Data File : VN087075.D
Acq On : 18 Jun 2025 11:01
Operator : JC\MD
Sample : VN0618MBL01
Misc : 5.00g/10mL/100uL/5.0mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0618MBL01

Quant Time: Jun 19 01:23:57 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 07 02:12:50 2025
Response via : Initial Calibration

