

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061825\
 Data File : VN087085.D
 Acq On : 18 Jun 2025 15:02
 Operator : JC\MD
 Sample : Q2347-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-10

Quant Time: Jun 19 01:29:42 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 02:12:50 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	247020	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	505646	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	456673	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	214657	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	178724	54.039	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	108.080%
35) Dibromofluoromethane	8.177	113	153679	51.285	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.560%
50) Toluene-d8	10.571	98	625267	52.709	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	105.420%
62) 4-Bromofluorobenzene	12.847	95	215540	48.905	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	97.800%
Target Compounds						
					Qvalue	
16) Acetone	4.453	43	44327	25.273	ug/l	95
18) Methyl Acetate	5.053	43	24876	4.867	ug/l	97
20) Methylene Chloride	5.294	84	9967	3.035	ug/l	92
37) Ethyl Acetate	7.577	43	17049	2.999	ug/l	99
43) Isopropyl Acetate	8.694	43	10704	1.173	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061825\
Data File : VN087085.D
Acq On : 18 Jun 2025 15:02
Operator : JC\MD
Sample : Q2347-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-10

Quant Time: Jun 19 01:29:42 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 07 02:12:50 2025
Response via : Initial Calibration

