

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061825\
 Data File : VN087088.D
 Acq On : 18 Jun 2025 16:08
 Operator : JC\MD
 Sample : Q2329-09
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TT149S1-20250613

Quant Time: Jun 19 01:30:47 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 02:12:50 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	253025	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	501869	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	459032	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	210865	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	180231	53.201	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 106.400%		
35) Dibromofluoromethane	8.177	113	157046	52.803	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 105.600%		
50) Toluene-d8	10.571	98	619300	52.599	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 105.200%		
62) 4-Bromofluorobenzene	12.847	95	221136	50.552	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 101.100%		
Target Compounds						
						Qvalue
19) Methyl tert-butyl Ether	5.818	73	57929	5.681	ug/l	98
44) Trichloroethene	9.353	130	9488	2.761	ug/l	98
64) Tetrachloroethene	11.106	164	2065	0.711	ug/l #	75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061825\
Data File : VN087088.D
Acq On : 18 Jun 2025 16:08
Operator : JC\MD
Sample : Q2329-09
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TT149S1-20250613

Quant Time: Jun 19 01:30:47 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 07 02:12:50 2025
Response via : Initial Calibration

