

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061825\
 Data File : VN087090.D
 Acq On : 18 Jun 2025 16:52
 Operator : JC\MD
 Sample : Q2329-14
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DUP04-20250613

Quant Time: Jun 19 01:31:27 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 02:12:50 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	187618	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	379825	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	344732	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	162129	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.589	65	137942	54.913	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 109.820%		
35) Dibromofluoromethane	8.177	113	118477	52.635	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 105.260%		
50) Toluene-d8	10.571	98	468161	52.538	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 105.080%		
62) 4-Bromofluorobenzene	12.847	95	162968	49.226	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 98.460%		
Target Compounds						
					Qvalue	
19) Methyl tert-butyl Ether	5.818	73	56006	7.407	ug/l	97
44) Trichloroethene	9.359	130	9162	3.523	ug/l	99
64) Tetrachloroethene	11.100	164	2049	0.939	ug/l #	77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061825\
Data File : VN087090.D
Acq On : 18 Jun 2025 16:52
Operator : JC\MD
Sample : Q2329-14
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DUP04-20250613

Quant Time: Jun 19 01:31:27 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 07 02:12:50 2025
Response via : Initial Calibration

