

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061825\
 Data File : VN087093.D
 Acq On : 18 Jun 2025 17:57
 Operator : JC\MD
 Sample : Q2329-17
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TB02-20250612

Quant Time: Jun 19 01:32:30 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 02:12:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	179779	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	370663	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	331725	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	154445	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	136079	56.534	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	113.060%
35) Dibromofluoromethane	8.177	113	114090	51.939	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.880%
50) Toluene-d8	10.571	98	453967	52.205	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	104.400%
62) 4-Bromofluorobenzene	12.847	95	158180	48.960	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	97.920%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061825\
Data File : VN087093.D
Acq On : 18 Jun 2025 17:57
Operator : JC\MD
Sample : Q2329-17
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TB02-20250612

Quant Time: Jun 19 01:32:30 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 07 02:12:50 2025
Response via : Initial Calibration

