

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061825\
 Data File : VN087095.D
 Acq On : 18 Jun 2025 18:39
 Operator : JC\MD
 Sample : Q2329-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FB01-20250613

Quant Time: Jun 19 01:33:09 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 02:12:50 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	177161	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	362021	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	325288	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	153113	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	133529	56.294	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	112.580%
35) Dibromofluoromethane	8.177	113	113292	52.806	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.620%
50) Toluene-d8	10.570	98	447665	52.709	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	105.420%
62) 4-Bromofluorobenzene	12.847	95	151423	47.988	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	95.980%
Target Compounds						Qvalue
16) Acetone	4.453	43	13976	11.111	ug/l	90
20) Methylene Chloride	5.300	84	4663	1.980	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061825\
Data File : VN087095.D
Acq On : 18 Jun 2025 18:39
Operator : JC\MD
Sample : Q2329-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FB01-20250613

Quant Time: Jun 19 01:33:09 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 07 02:12:50 2025
Response via : Initial Calibration

