

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061919\
 Data File : VN056383.D
 Acq On : 19 Jun 2019 13:46
 Operator : JC/SP
 Sample : K3441-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SAMPLE-3

Quant Time: Jun 21 04:09:10 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061919W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 21 03:45:40 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	375890	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	636492	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	640515	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	194023	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	212847	51.09	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.18%	
35) Dibromofluoromethane	7.59	113	184932	46.57	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.14%	
50) Toluene-d8	10.09	98	802538	51.94	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.88%	
62) 4-Bromofluorobenzene	12.40	95	266279	51.70	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.40%	
Target Compounds						
16) Acetone	3.82	43	16716	8.009	ug/l	93
20) Methylene Chloride	4.55	84	138717	34.987	ug/l	98
43) Isopropyl Acetate	8.17	43	12993	1.530	ug/l #	90
44) Trichloroethene	8.84	130	7891	1.700	ug/l	99
68) m/p-Xylenes	11.62	106	9222	1.114	ug/l	87
95) Naphthalene	15.13	128	8021	6.053	ug/l #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN061919\
Data File : VN056383.D
Acq On : 19 Jun 2019 13:46
Operator : JC/SP
Sample : K3441-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SAMPLE-3

Quant Time: Jun 21 04:09:10 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061919W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 21 03:45:40 2019
Response via : Initial Calibration

