


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN061919\
Data File : VN056371.D
Acq On : 19 Jun 2019 8:23
Operator : JC/SP
Sample : BFB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061919W.M
Title : SW846 8260
Last Update : Fri Jun 21 03:45:40 2019



AutoFind: Scans 3304, 3305, 3306; Background Corrected with Scan 3291

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.2	9024	PASS
75	95	30	60	44.7	23413	PASS
95	95	100	100	100.0	52349	PASS
96	95	5	9	7.1	3695	PASS
173	174	0.00	2	0.7	318	PASS
174	95	50	100	89.9	47050	PASS
175	174	5	9	7.4	3500	PASS
176	174	95	101	96.8	45554	PASS
177	176	5	9	6.5	2955	PASS