

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061920\
 Data File : VN061945.D
 Acq On : 19 Jun 2020 12:12
 Operator : JC/MD
 Sample : VN0619WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0619WBL01

Quant Time: Jun 20 07:53:09 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 17 01:19:44 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	147593	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	270285	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	260026	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	101714	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	99067	50.05	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.10%	
35) Dibromofluoromethane	7.55	113	86129	49.81	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.62%	
50) Toluene-d8	10.06	98	337363	49.63	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.26%	
62) 4-Bromofluorobenzene	12.38	95	111388	47.77	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.54%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN061920\
Data File : VN061945.D
Acq On : 19 Jun 2020 12:12
Operator : JC/MD
Sample : VN0619WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0619WBL01

Quant Time: Jun 20 07:53:09 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061620W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 17 01:19:44 2020
Response via : Initial Calibration

