

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061920\
 Data File : VN061977.D
 Acq On : 20 Jun 2020 5:15
 Operator : JC/MD
 Sample : L3043-05
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW203D1-20200614

Quant Time: Jun 23 12:03:08 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 17 01:19:44 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	138959	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	261052	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	247999	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	89335	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	96721	51.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.80%	
35) Dibromofluoromethane	7.55	113	84413	50.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.08%	
50) Toluene-d8	10.06	98	327831	49.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.86%	
62) 4-Bromofluorobenzene	12.38	95	104390	46.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.70%	
Target Compounds						
12) 1,1-Dichloroethene	3.70	96	8431	5.971	ug/l	96
24) 1,1-Dichloroethane	5.80	63	60860	21.693	ug/l	98
27) cis-1,2-Dichloroethene	6.78	96	1693	0.905	ug/l	76
30) Chloroform	7.33	83	3440	1.147	ug/l	100
32) 1,1,1-Trichloroethane	7.53	97	8430	3.301	ug/l	96
44) Trichloroethene	8.80	130	13065	6.883	ug/l	96
64) Tetrachloroethene	10.60	164	1833	1.160	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN061920\
Data File : VN061977.D
Acq On : 20 Jun 2020 5:15
Operator : JC/MD
Sample : L3043-05
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW203D1-20200614

Quant Time: Jun 23 12:03:08 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061620W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 17 01:19:44 2020
Response via : Initial Calibration

