

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061920\
 Data File : VN061979.D
 Acq On : 20 Jun 2020 6:04
 Operator : JC/MD
 Sample : L3043-09
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TT101D-20200616

Quant Time: Jun 23 12:04:32 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 17 01:19:44 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	133909	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	245739	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	230708	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	83675	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	92812	51.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.36%	
35) Dibromofluoromethane	7.55	113	79671	50.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.36%	
50) Toluene-d8	10.06	98	309920	50.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.28%	
62) 4-Bromofluorobenzene	12.38	95	95783	45.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.36%	
Target Compounds						
2) Dichlorodifluoromethane	1.83	85	2014	1.441	ug/l	99
9) 1,1,2-Trichlorotrifluoroet	3.72	101	22423	16.513	ug/l	99
12) 1,1-Dichloroethene	3.70	96	4497	3.305	ug/l	83
27) cis-1,2-Dichloroethene	6.78	96	6542	3.628	ug/l	91
30) Chloroform	7.33	83	1695	0.587	ug/l	92
44) Trichloroethene	8.80	130	194422	108.805	ug/l	99
64) Tetrachloroethene	10.61	164	1135	0.772	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN061920\
Data File : VN061979.D
Acq On : 20 Jun 2020 6:04
Operator : JC/MD
Sample : L3043-09
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TT101D-20200616

Quant Time: Jun 23 12:04:32 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061620W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 17 01:19:44 2020
Response via : Initial Calibration

