

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061920\
 Data File : VN061980.D
 Acq On : 20 Jun 2020 6:28
 Operator : JC/MD
 Sample : L3043-13
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW203D2-20200613

Quant Time: Jun 23 12:05:06 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 17 01:19:44 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	135076	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	245243	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	235016	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	84007	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	90663	50.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.10%	
35) Dibromofluoromethane	7.55	113	79125	50.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.86%	
50) Toluene-d8	10.06	98	308931	50.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.18%	
62) 4-Bromofluorobenzene	12.38	95	97006	45.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.70%	
Target Compounds						
16) Acetone	3.79	43	3565	6.662	ug/l	89
44) Trichloroethene	8.81	130	1873	1.050	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN061920\
Data File : VN061980.D
Acq On : 20 Jun 2020 6:28
Operator : JC/MD
Sample : L3043-13
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW203D2-20200613

Quant Time: Jun 23 12:05:06 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061620W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 17 01:19:44 2020
Response via : Initial Calibration

