

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061920\
 Data File : VN061982.D
 Acq On : 20 Jun 2020 7:16
 Operator : JC/MD
 Sample : L3043-15
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW202D-20200613

Quant Time: Jun 23 12:06:11 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 17 01:19:44 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	130814	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	242604	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	231305	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	83430	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	91605	52.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.42%	
35) Dibromofluoromethane	7.55	113	78758	50.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.48%	
50) Toluene-d8	10.06	98	308025	50.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.96%	
62) 4-Bromofluorobenzene	12.38	95	97144	46.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.82%	
Target Compounds						
					Qvalue	
7) Trichlorofluoromethane	2.98	101	6118	2.366	ug/l	100
12) 1,1-Dichloroethene	3.70	96	2146	1.615	ug/l #	76
16) Acetone	3.78	43	2791	5.386	ug/l	98
27) cis-1,2-Dichloroethene	6.78	96	24505	13.911	ug/l	89
30) Chloroform	7.33	83	1845	0.654	ug/l	87
32) 1,1,1-Trichloroethane	7.52	97	1015	0.422	ug/l #	51
44) Trichloroethene	8.80	130	133595	75.730	ug/l	97
64) Tetrachloroethene	10.60	164	26828	18.198	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN061920\
Data File : VN061982.D
Acq On : 20 Jun 2020 7:16
Operator : JC/MD
Sample : L3043-15
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW202D-20200613

Quant Time: Jun 23 12:06:11 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061620W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 17 01:19:44 2020
Response via : Initial Calibration

