

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061920\
 Data File : VN061983.D
 Acq On : 20 Jun 2020 7:40
 Operator : JC/MD
 Sample : L3043-16
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW201D-20200613

Quant Time: Jun 22 03:31:29 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 17 01:19:44 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	127160	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	235344	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	229467	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	80568	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	7.99	65	90053	52.80	ug/l	0.00
Spiked Amount			Recovery	=	105.60%	
35) Dibromofluoromethane	7.55	113	76876	51.06	ug/l	0.00
Spiked Amount			Recovery	=	102.12%	
50) Toluene-d8	10.06	98	297956	50.34	ug/l	0.00
Spiked Amount			Recovery	=	100.68%	
62) 4-Bromofluorobenzene	12.38	95	94401	46.49	ug/l	0.00
Spiked Amount			Recovery	=	92.98%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) Dichlorodifluoromethane	1.83	85	15931	9.864	ug/l	99
4) Vinyl Chloride	2.16	62	1143	0.543	ug/l	95
7) Trichlorofluoromethane	2.98	101	489625	194.807	ug/l	98
9) 1,1,2-Trichlorotrifluoroet	3.72	101	4885	3.788	ug/l #	82
12) 1,1-Dichloroethene	3.70	96	73446	56.847	ug/l	97
16) Acetone	3.78	43	3235	6.422	ug/l	93
24) 1,1-Dichloroethane	5.81	63	12527	4.880	ug/l	96
27) cis-1,2-Dichloroethene	6.78	96	5952	3.476	ug/l	87
30) Chloroform	7.33	83	15005	5.469	ug/l	95
32) 1,1,1-Trichloroethane	7.53	97	19294	8.257	ug/l	98
44) Trichloroethene	8.80	130	193374	112.998	ug/l	99
64) Tetrachloroethene	10.60	164	172500	117.945	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN061920\
Data File : VN061983.D
Acq On : 20 Jun 2020 7:40
Operator : JC/MD
Sample : L3043-16
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW201D-20200613

Quant Time: Jun 22 03:31:29 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061620W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 17 01:19:44 2020
Response via : Initial Calibration

