

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061920\
 Data File : VN061984.D
 Acq On : 20 Jun 2020 8:04
 Operator : JC/MD
 Sample : L3043-20
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE139D1-20200612

Quant Time: Jun 23 12:07:29 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 17 01:19:44 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	136333	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	253024	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	245060	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	92378	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	95043	51.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.96%	
35) Dibromofluoromethane	7.55	113	81900	50.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.20%	
50) Toluene-d8	10.06	98	320389	50.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.70%	
62) 4-Bromofluorobenzene	12.38	95	103218	47.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.56%	
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluoroet	3.72	101	4922	3.560	ug/l	97
12) 1,1-Dichloroethene	3.70	96	3501	2.527	ug/l	99
16) Acetone	3.79	43	3381	6.260	ug/l #	82
27) cis-1,2-Dichloroethene	6.78	96	2550	1.389	ug/l	84
30) Chloroform	7.33	83	2135	0.726	ug/l #	83
38) Carbon Tetrachloride	7.73	117	2160	0.945	ug/l	89
44) Trichloroethene	8.80	130	229389	124.677	ug/l	99
55) 1,1,2-Trichloroethane	10.53	97	1505	0.871	ug/l #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN061920\
Data File : VN061984.D
Acq On : 20 Jun 2020 8:04
Operator : JC/MD
Sample : L3043-20
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE139D1-20200612

Quant Time: Jun 23 12:07:29 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061620W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 17 01:19:44 2020
Response via : Initial Calibration

