

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061920\
 Data File : VN061986.D
 Acq On : 20 Jun 2020 8:52
 Operator : JC/MD
 Sample : L3043-23
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 REDUP06-20200616

Quant Time: Jun 23 12:09:09 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 17 01:19:44 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	136311	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	249984	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	246794	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	93185	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	7.99	65	95307	52.13	ug/l	0.00
Spiked Amount			Recovery	=	104.26%	
35) Dibromofluoromethane	7.55	113	81189	50.77	ug/l	0.00
Spiked Amount			Recovery	=	101.54%	
50) Toluene-d8	10.06	98	319385	50.80	ug/l	0.00
Spiked Amount			Recovery	=	101.60%	
62) 4-Bromofluorobenzene	12.38	95	104662	48.53	ug/l	0.00
Spiked Amount			Recovery	=	97.06%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
9) 1,1,2-Trichlorotrifluoroet	3.72	101	7491	5.419	ug/l	97
12) 1,1-Dichloroethene	3.70	96	4123	2.977	ug/l	92
16) Acetone	3.79	43	3026	5.604	ug/l #	86
27) cis-1,2-Dichloroethene	6.79	96	3222	1.755	ug/l	58
30) Chloroform	7.33	83	1300	0.442	ug/l	92
44) Trichloroethene	8.80	130	1098490	604.310	ug/l	99
64) Tetrachloroethene	10.60	164	9274	5.896	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN061920\
Data File : VN061986.D
Acq On : 20 Jun 2020 8:52
Operator : JC/MD
Sample : L3043-23
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
REDUP06-20200616

Quant Time: Jun 23 12:09:09 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061620W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 17 01:19:44 2020
Response via : Initial Calibration

