


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN061920\
 Data File : VN061968.D
 Acq On : 19 Jun 2020 22:00
 Operator : JC/MD
 Sample : BFB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061620W.M
 Title : SW846 8260
 Last Update : Wed Jun 17 01:19:44 2020



AutoFind: Scans 3297, 3298, 3299; Background Corrected with Scan 3284

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.5	7474	PASS
75	95	30	60	50.8	20570	PASS
95	95	100	100	100.0	40501	PASS
96	95	5	9	7.2	2906	PASS
173	174	0.00	2	1.8	569	PASS
174	95	50	100	78.9	31965	PASS
175	174	5	9	8.0	2545	PASS
176	174	95	101	97.3	31109	PASS
177	176	5	9	6.6	2053	PASS