

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061923\
 Data File : VN078271.D
 Acq On : 19 Jun 2023 11:37
 Operator : JC\MD
 Sample : VN0619MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0619MBL01

Quant Time: Jun 19 16:02:29 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 07 05:05:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	427697	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	782070	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	696489	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	217128	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	304319	53.366	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 106.740%	
35) Dibromofluoromethane	8.177	113	275844	55.783	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 111.560%	
50) Toluene-d8	10.570	98	908860	47.676	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 95.360%	
62) 4-Bromofluorobenzene	12.859	95	283189	47.785	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 95.580%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061923\
Data File : VN078271.D
Acq On : 19 Jun 2023 11:37
Operator : JC\MD
Sample : VN0619MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0619MBL01

Quant Time: Jun 19 16:02:29 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 07 05:05:00 2023
Response via : Initial Calibration

