

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061923\
 Data File : VN078276.D
 Acq On : 19 Jun 2023 13:48
 Operator : JC\MD
 Sample : 03260-01
 Misc : 5.02g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-1

Quant Time: Jun 19 16:03:28 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 07 05:05:00 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	371424	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	681912	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	596244	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	209047	50.000	ug/l	# 0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	254557	51.403	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 102.800%		
35) Dibromofluoromethane	8.171	113	222832	51.681	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 103.360%		
50) Toluene-d8	10.571	98	826635	49.732	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 99.460%		
62) 4-Bromofluorobenzene	12.853	95	249817	48.346	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 96.700%		
Target Compounds						
						Qvalue
18) Methyl Acetate	5.036	43	14526	2.213	ug/l	# 86
84) 1,2,4-Trimethylbenzene	13.488	105	22998	1.482	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061923\
Data File : VN078276.D
Acq On : 19 Jun 2023 13:48
Operator : JC\MD
Sample : 03260-01
Misc : 5.02g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-1

Quant Time: Jun 19 16:03:28 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 07 05:05:00 2023
Response via : Initial Calibration

