

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061923\
 Data File : VN078277.D
 Acq On : 19 Jun 2023 14:12
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Jun 20 01:27:07 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 07 05:05:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	349797	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	653636	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	563087	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	155094	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	253939	54.448	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	108.900%
35) Dibromofluoromethane	8.177	113	218472	52.862	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.720%
50) Toluene-d8	10.577	98	793186	49.784	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.560%
62) 4-Bromofluorobenzene	12.853	95	223089	45.041	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.080%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061923\
Data File : VN078277.D
Acq On : 19 Jun 2023 14:12
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Jun 20 01:27:07 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 07 05:05:00 2023
Response via : Initial Calibration

