

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061923\
 Data File : VN078279.D
 Acq On : 19 Jun 2023 15:00
 Operator : JC\MD
 Sample : 03310-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB-RW11-GW-898-900

Quant Time: Jun 20 01:28:10 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 07 05:05:00 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	484272	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	879572	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	779127	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	249763	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	342466	53.039	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 106.080%	
35) Dibromofluoromethane	8.177	113	303974	54.658	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 109.320%	
50) Toluene-d8	10.571	98	1032371	48.152	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 96.300%	
62) 4-Bromofluorobenzene	12.853	95	335376	50.318	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 100.640%	
Target Compounds						
					Qvalue	
16) Acetone	4.453	43	6963	3.747	ug/l #	79
17) Carbon Disulfide	4.724	76	7078	0.520	ug/l	100
93) 1,2,4-Trichlorobenzene	15.406	180	2503	0.775	ug/l	84
95) Naphthalene	15.653	128	7703	0.709	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061923\
Data File : VN078279.D
Acq On : 19 Jun 2023 15:00
Operator : JC\MD
Sample : 03310-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-RW11-GW-898-900

Quant Time: Jun 20 01:28:10 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 07 05:05:00 2023
Response via : Initial Calibration

